



**Abbott**

# **Predicted Impact of Variants on Abbott's SARS-CoV-2/COVID-19 Diagnostic Tests**

Technical Brief

January 31, 2023

**Purpose:** This Technical Brief is an up-to-date overview on the predicted impact, if any, to the performance of Abbott's SARS-CoV-2/COVID-19 diagnostic tests in the detection of SARS-CoV-2 viral variants, as determined through ongoing analysis by the Abbott Pandemic Defense Coalition. This document is provided as assurance to customers that Abbott is conducting continuous and thorough analysis of emerging SARS-CoV-2 variants.

**Background:** Emerging variants of the SARS-CoV-2 virus have been identified across the globe with concerning pathogenic properties.<sup>1,2</sup> Assessing the risk emerging variants may pose to public health relies on continued identification and characterization.<sup>3</sup> Concerns have been raised as some variants have been reported to have increased viral transmission and disease severity.<sup>4</sup> As these variants are identified, it is imperative that efforts are taken to monitor any potential impact the genomic mutations have on viral detection by Abbott's diagnostic tests.

**Abbott's Monitoring:** Abbott is continuously monitoring the global SARS-CoV-2 situation through complex processes overseen by the Abbott Pandemic Defense Coalition.<sup>5,6,7</sup> As emerging variants are identified, sequence and *in silico* analyses are conducted to evaluate potential impact of these mutations to our tests. This proactive monitoring scheme enables Abbott to communicate the most up to date information specific to our tests. While the detailed evidence is proprietary, Abbott recognizes the need to provide customer assurance on our test performance. In addition to this document, the Abbott Pandemic Defense Coalition has published a study evaluating Abbott's molecular, antigen, and serologic assays with several SARS-CoV-2 viral variants and will continue to publish as evaluations of emerging variants continue to arise.<sup>6</sup>

**Predicted Impact of Variants on Abbott's SARS-CoV-2/COVID-19 Diagnostic Tests:** The following table (Table 1) lists the Abbott's SARS-CoV-2/COVID-19 diagnostic tests, the target(s) detected, and any predicted impact on assay performance based on data analyses to date (see **Table 2, Summary of Variants Analyzed to Date**).

**Table 1: Predicted Impact of Variants on Abbott’s SARS-CoV-2/COVID-19 Diagnostic Tests:**

<b>Abbott’s SARS-CoV-2/COVID-19 Test</b>	<b>SARS-CoV-2 Detected Target(s)</b>	<b>Test Performance</b>
Panbio™ COVID-19 Ag Rapid Test Device	N* protein	No Predicted Impact
Panbio™ COVID-19 Antigen Self-Test	N protein	No Predicted Impact
Panbio™ COVID-19 IgG/IgM Rapid Test Device	N protein	No Predicted Impact
Panbio™ COVID-19/Flu A&B Rapid Panel	N protein	No Predicted Impact
BinaxNOW™ COVID-19 Ag Card	N protein	No Predicted Impact
BinaxNOW™ COVID-19 Antigen Self Test	N protein	No Predicted Impact
ID NOW™ COVID-19 Test	RdRp** gene	No Predicted Impact
ID NOW™ COVID-19 2.0 Test	RdRp gene	No Predicted Impact
Alinity m SARS-CoV-2	RdRp and N genes	No Predicted Impact
Alinity m Resp-4-Plex	RdRp and N genes	No Predicted Impact
RealTime SARS-CoV-2	RdRp and N genes	No Predicted Impact
Alinity i SARS-CoV-2 IgG	N protein	No Predicted Impact
ARCHITECT SARS-CoV-2 IgG	N protein	No Predicted Impact
Alinity i SARS-CoV-2 IgM	S*** protein	No Predicted Impact
ARCHITECT SARS-CoV-2 IgM	S protein	No Predicted Impact
AdviseDx SARS-CoV-2 IgM (Alinity i)	S protein	No Predicted Impact
AdviseDx SARS-CoV-2 IgM (ARCHITECT)	S protein	No Predicted Impact
Alinity i SARS-CoV-2 IgG II Quant	S protein	No Predicted Impact
ARCHITECT SARS-CoV-2 IgG II Quant	S protein	No Predicted Impact
AdviseDx SARS-CoV-2 IgG II (Alinity i)	S protein	No Predicted Impact
AdviseDx SARS-CoV-2 IgG II (ARCHITECT)	S protein	No Predicted Impact

\*N – Nucleocapsid; \*\*RdRp – RNA dependent RNA polymerase; \*\*\* S - Spike

**Table 2: Summary of Variants Analyzed to Date:**<sup>2-4,6,8,9,10</sup>

<b>WHO Nomenclature</b>	<b>Lineage</b>
Alpha <sup>#</sup>	B.1.1.7
Alpha <sup>#</sup>	Q.5
Alpha <sup>#</sup>	Q.6
Alpha <sup>#</sup>	Q.7
Beta	B.1.351
Beta	B.1.351.2
Beta	B.1.351.3
Beta	B.1.351.5
Gamma	P.1
Gamma	P.1.1
Gamma	P.1.2
Delta <sup>*</sup>	B.1.617.2
Delta <sup>*</sup>	AY.1
Delta <sup>*</sup>	AY.2
Delta <sup>*</sup>	AY.3
Delta <sup>*</sup>	AY.3.1
Delta <sup>*</sup>	AY.4
Delta <sup>*</sup>	AY.4.2
Delta <sup>*</sup>	AY.5
Delta <sup>*</sup>	AY.5.1
Delta <sup>*</sup>	AY.5.2
Delta <sup>*</sup>	AY.6
Delta <sup>*</sup>	AY.7
Delta <sup>*</sup>	AY.8
Delta <sup>*</sup>	AY.9

Delta*	AY.10
Delta*	AY.11
Delta*	AY.12
Delta*	AY.25
Delta*	AY.27
Delta*	AY.30
Delta*	AY.31
Delta*	AY.70
Delta*	AY.74
Delta*	AY.88
Delta*	AY.97
Delta*	AY.107
Delta*/Omicron^&	XD@
Delta*/Omicron^&	XF
Epsilon	B.1.427
Epsilon	B.1.429
Zeta	P.2
Eta	B.1.525
Theta	P.3
Iota	B.1.526
Kappa	B.1.617.1
Lambda	C.37
Mu	B.1.621
Mu	B.1.621.1
Omicron^	B.1.1.529
Omicron^	BA.1
Omicron^	BA.1.1

Omicron^	BA.1.15
Omicron^	BA.2
Omicron^	BA.2.2
Omicron^	BA.2.3
Omicron^	BA.2.9
Omicron^	BA.2.9.1
Omicron^	BA.2.10
Omicron^	BA.2.12
Omicron^	BA.2.12.1
Omicron^	BA.2.16
Omicron^	BA.2.3.20
Omicron^	BA.2.38
Omicron^	BA.2.38.1
Omicron^	BA.2.75
Omicron^	BA.2.75.2
Omicron^	BA.2.76
Omicron^	BA.3
Omicron^	BA.4
Omicron^	BA.4.1
Omicron^	BA.4.6
Omicron^	BA.4.7
Omicron^	BA.5
Omicron^	BA.5.1
Omicron^	BA.5.1.12
Omicron^	BA.5.1.25
Omicron^	BA.5.2
Omicron^	BA.5.2.1

Omicron^	BA.5.3
Omicron^	BA.5.3.1
Omicron^	BA.5.3.5
Omicron^	BA.5.5
Omicron^	BA.5.6
Omicron^	BE.1
Omicron^	BE.1.1
Omicron^	BF.5
Omicron^	BF.7
Omicron^	BF.7.4
Omicron^~	BQ.1
Omicron^~	BQ.1.1
Omicron^~	BQ.1.1.13
Omicron^~	BQ.1.1.18
Omicron^~	BQ.1.1.20
Omicron^~	BQ.1.1.22
Omicron^~	BQ.1.8
Omicron^	BN.1
Omicron^	BW.1
Omicron^	CH.1.1
Omicron^	CK.2.1.1
Omicron^%	XE
Omicron^~	XBB
Omicron^~	XBB.1
Omicron^~	XBB.1.5
Not designated	A.23.1+E484K
Not designated	A.27

Not designated	AT.1
Not designated	AV.1
Not designated	B.1.1.318
Not designated	B.1.1.451
Not designated	B.1.1.519
Not designated	B.1.1.523
Not designated	B.1.1.7 with E484K
Not designated	B.1.214.2
Not designated	B.1.36.26
Not designated	B.1.429.1
Not designated	B.1.466.2
Not designated	B.1.526.1
Not designated	B.1.526.2
Not designated	B.1.616
Not designated	B.1.617.3
Not designated	B.1.618
Not designated	B.1.619
Not designated	B.1.620
Not designated	B.1.628
Not designated	C.1.2
Not designated	C.36.3
Not designated	C.36.3.1
Not designated	R.1
Not designated	P.4

# Includes all Q lineages, which as noted by the WHO, is an alias for B.1.1.7 in Pango nomenclature.<sup>9,10</sup>

\* Includes all AY lineages, which as noted by the WHO, is an alias for B.1.617.2 in Pango nomenclature.<sup>11</sup>

^ Includes all BA and BE lineages, which is an alias for B.1.1.529 in Pango nomenclature.<sup>12, 13, 15, 16</sup>

& XD and XF are recombinant variants of Delta and Omicron BA.1.<sup>14</sup>

% XE is a recombinant variant of Omicron BA.1 and BA.2.<sup>14</sup>

@ *In silico* analysis of the XD variant identified the presence of a mutation in one of the Abbott test targets. This mutation is found in the N gene of the Delta variant and was shown to have no impact.<sup>8</sup>

~WHO's Technical Advisory group on SARS-CoV-2 Virus Evolution (TAG-VE) has decided that sublineages XBB and BQ.1 remain a part of the Omicron variant.<sup>17</sup>



## Technical Support:

If you have any questions on the provided information or are able to provide access to emerging variant samples, please contact Technical Support.

### ID NOW™ COVID-19 test and ID NOW™ COVID-19 2.0 test<sup>1</sup>:

<https://www.globalpointofcare.abbott/en/product-details/id-now-covid-19.html>

### BinaxNOW™ COVID-19 Ag Card<sup>2</sup>:

Professional: <https://www.globalpointofcare.abbott/en/product-details/navica-binaxnow-covid-19-us.html>

Proctored: <https://www.globalpointofcare.abbott/en/product-details/binaxnow-covid-19-home-test-us.html>

### BinaxNOW™ COVID-19 Self Test:

<https://www.globalpointofcare.abbott/en/product-details/binaxnow-covid-19-antigen-self-test-us.html>

### Panbio™ COVID-19 Ag Rapid Test Device<sup>3</sup>:

<https://www.globalpointofcare.abbott/en/product-details/panbio-covid-19-ag-antigen-test.html>

### Panbio™ COVID-19 IgG/IgM Rapid Test Device<sup>4</sup>:

<https://www.globalpointofcare.abbott/en/product-details/panbio-covid-19-igg-igm-antibody-test.html>

### Panbio™ COVID-19 Antigen Self-Test<sup>5</sup>:

<https://www.globalpointofcare.abbott/en/product-details/panbio-covid-19-antigen-self-test.html>

### Panbio™ COVID-19/Flu A&B Rapid Panel<sup>6</sup>:

<https://www.globalpointofcare.abbott/en/product-details/panbio-covid-19-flu-ab-rapid-panel.html>

### Alinity m SARS-CoV-2<sup>7</sup>, Alinity m Resp-4-Plex<sup>8</sup>, Abbott RealTime SARS-CoV-2<sup>8</sup>:

Global: <https://www.molecular.abbott/int/en/contact-technical-support>

US: <https://www.molecular.abbott/us/en/knowledge-center/support>

### Abbott's SARS-CoV-2 IgM, SARS-CoV-2 IgG, SARS-CoV-2 IgG II Quant, AdviseDx SARS-CoV-2 IgM, and AdviseDx SARS-CoV-2 IgG II Assays for the Use with ARCHITECT and Alinity i<sup>8</sup>:

<https://www.corelaboratory.abbott/int/en/about-us/customer-service-support>

---

<sup>1</sup> CDC. Variants of the Virus. Updated August 11, 2021. Accessed February 1, 2023. <https://www.cdc.gov/coronavirus/2019-ncov/variants/index.html>

<sup>2</sup> UK Health Security Agency. Research and analysis. Variants: distribution of cases data, 20 May 2021. Updated June 24, 2022. Accessed February 1, 2023. <https://www.gov.uk/government/publications/covid-19-variants-genomically-confirmed-case-numbers/variants-distribution-of-cases-data>

<sup>3</sup> CDC. SARS-CoV-2 Variant Classifications and Definitions. Updated April 26, 2022. Accessed February 1, 2023. <https://www.cdc.gov/coronavirus/2019-ncov/variants/variant-info.html>

<sup>4</sup> ECDC. SARS-CoV-2 variants of concern as of 8 December 2022. Updated December 14, 2022. Accessed February 1, 2023. <https://www.ecdc.europa.eu/en/covid-19/variants-concern>

<sup>5</sup> Abbott Newsroom. How We're Tracking COVID-19 Variants. Updated February 23, 2021. Accessed February 1, 2023. <https://www.abbott.com/corpnnewsroom/products-and-innovation/how-we-track-covid-19-variants.html>

<sup>6</sup> Rodgers MA, Olivo A, Harris BJ, *et al.* Detection of SARS-CoV-2 variants by Abbott molecular, antigen, and serological tests. *J Clin Virol.* 2022;147:105080. <https://doi.org/10.1016/j.jcv.2022.105080>

<sup>7</sup> Averhoff F, Berg M, Rodgers M, *et al.* The Abbott Pandemic Defense Coalition: a unique multisector approach adds to global pandemic preparedness efforts [published online ahead of print, 2022 Feb 5]. *Int J Infect Dis.* 2022;117:356-360. doi: <https://doi.org/10.1016/j.ijid.2022.02.001>

<sup>8</sup> Abbott data on file

- 
- <sup>9</sup> WHO. Coronavirus disease (COVID-19) Weekly Epidemiological Update and Weekly Operational Update. Accessed February 1, 2023. <https://www.who.int/emergencies/diseases/novel-coronavirus-2019/situation-reports>
- <sup>10</sup> WHO. Tracking SARS-CoV-2 Variants. Updated January 26, 2023. Accessed February 1, 2023. <https://www.who.int/en/activities/tracking-SARS-CoV-2-variants/>
- <sup>11</sup> Pango Network. New AY Lineages. Updated August 13, 2021. Accessed February 1, 2023. <http://pango.network/new-ay-lineages/>
- <sup>12</sup> Pango Network. Updates to Omicron Lineage B.1.1.529. Updated December 9, 2021. Accessed February 1, 2023. <https://www.pango.network/updates-to-omicron-lineage-b-1-1-529/>
- <sup>13</sup> WHO. Statement on Omicron sublineage BA.2. Updated February 22, 2022. Accessed February 1, 2023. <https://www.who.int/news/item/22-02-2022-statement-on-omicron-sublineage-ba.2>
- <sup>14</sup> UK Security Agency. SARS-CoV-2 variants of concern and variants under investigation in England. Updated March 25, 2022. Accessed February 1, 2023. [https://assets.publishing.service.gov.uk/government/uploads/system/uploads/attachment\\_data/file/1063424/Tech-Briefing-39-25March2022\\_FINAL.pdf](https://assets.publishing.service.gov.uk/government/uploads/system/uploads/attachment_data/file/1063424/Tech-Briefing-39-25March2022_FINAL.pdf)
- <sup>15</sup> Pango Network. Summary of Designated Omicron Lineages. Updated April 1, 2022. Accessed February 1, 2023. <https://www.pango.network/summary-of-designated-omicron-lineages/>
- <sup>16</sup> Cov-Lineages. Lineage List. Accessed February 1, 2023. [https://cov-lineages.org/lineage\\_list.html](https://cov-lineages.org/lineage_list.html)
- <sup>17</sup> CDC. TAG-VE statement on Omicron sublineages BQ.1 and XBB. 27 Oct 2022. Accessed February 1, 2023. <https://www.who.int/news/item/27-10-2022-tag-ve-statement-on-omicron-sublineages-bq.1-and-xbb>

#### FOR EXTERNAL USE

Products not available in all countries. Available to consumers in select markets.

#The Panbio™ COVID-19 Ag Rapid Test Device, Panbio™ COVID-19 IgG/IgM Rapid Test Device, Panbio™ COVID-19 Antigen Self-Test, Panbio™ COVID-19/Flu A&B Rapid Panel, and SARS-CoV-2 IgM and SARS-CoV-2 IgG II Quant Assays for the Use with ARCHITECT and Alinity i are not available for sale in the US.

^ Emergency Use Authorization (EUA) Conditions for BinaxNOW™ COVID-19 Ag Card, BinaxNOW™ COVID-19 Antigen Self Test, ID NOW™ COVID-19, ID NOW™ COVID-19 2.0, Alinity m SARS-CoV-2, Alinity m Resp-4-Plex and Realtime SARS-CoV-2 assay, SARS-CoV-2 IgG, AdviseDx SARS-CoV-2 IgM, and AdviseDx SARS-CoV-2 IgG II assays for the use with ARCHITECT and Alinity i:

- BinaxNOW™ COVID-19 Ag Card has not been FDA cleared or approved, but have been authorized for emergency use by FDA under an EUA. It has been authorized only for the detection of proteins from SARS-CoV-2, not for any other viruses or pathogens;
- The BinaxNOW™ COVID-19 Antigen Self Test has not been FDA cleared or approved. It has been authorized by the FDA under an emergency use authorization. It has been authorized only for the detection of proteins from SARS-CoV-2, not for any other viruses or pathogens. BinaxNOW™ COVID-19 Antigen Self Test should be performed twice in 3 days, at least 24 hours apart (and no more than 48 hours) apart;
- ID NOW™ COVID-19 and ID NOW™ COVID-19 2.0 have not been FDA cleared or approved, but have been authorized for emergency use by FDA under an EUA for use by authorized laboratories or patient care settings;
- Alinity m SARS-CoV-2, Alinity m Resp-4-Plex and Realtime SARS-CoV-2 assays have not been FDA cleared or approved, but have been authorized for emergency use by FDA under an EUA for use by authorized laboratories;
- Alinity m SARS-CoV-2 and Alinity m Resp-4-Plex assays have been authorized by the FDA under an EUA for use by laboratories certified under CLIA, to perform moderate or high complexity tests;
- ID NOW™ COVID-19, ID NOW™ COVID-19 2.0, Alinity m SARS-CoV-2 assay and RealTime SARS-CoV-2 assay have been authorized only for the detection of nucleic acid from SARS-CoV-2, not for any other viruses or pathogens;
- Alinity m Resp-4-Plex has been authorized only for the detection and differentiation of nucleic acid from SARS-CoV-2, influenza A, influenza B, and/or Respiratory Syncytial Virus, not for any other viruses or pathogens;
- SARS-CoV-2 IgG, AdviseDx SARS-CoV-2 IgM, and AdviseDx SARS-CoV-2 IgG II Assays for the Use with ARCHITECT and Alinity i have not been FDA cleared or approved, but has been authorized for emergency use by FDA under an EUA for use by authorized laboratories. These products have been authorized only for detecting the presence of IgM or IgG antibodies against SARS-CoV-2, not for any other viruses or pathogens. Prescription Use Only.
- The emergency use of the products are only authorized for the duration of the declaration that circumstances exist justifying the authorization of emergency use of *in vitro* diagnostics for detection and/or diagnosis of COVID-19 under Section 564(b)(1) of the Federal Food, Drug and Cosmetic Act, 21 U.S.C. § 360bbb-3(b)(1), unless the declaration is terminated or authorization is revoked sooner.

© 2022 Abbott. All rights reserved. All trademarks referenced are trademarks of either the Abbott group of companies or their respective owners.