

Abbott RealTime IDH1 Control Kit

en

REF 08N90-080

51-608284/R2

REF 08N90-080
51-608284/R2

Key to Symbols Used	
	List Number
	In Vitro Diagnostic Medical Device
	Lot Number
	Authorized Representative in the European Community
	In Vitro Test
	For In Vitro Diagnostic Use
	Product of USA
	Prescription Use Only
	CE Mark
	Use by
	Negative Control
	Positive Control
	Temperature Limit
	Consult instructions for use
	Warning
	Manufacturer

CUSTOMER SERVICE: 1-800-553-7042

This package insert must be read carefully prior to use. Package insert instructions must be followed accordingly. Reliability of assay results cannot be guaranteed if there are any deviations from the instructions in this package insert.

NAME

Abbott RealTime IDH1 Control Kit

INTENDED USE

The Abbott RealTime IDH1 Controls are used to establish run validity of Abbott RealTime IDH1 when used for the qualitative detection of single nucleotide variants (SNVs) coding five IDH1 R132 mutations (R132C, R132H, R132G, R132S, and R132L) in DNA extracted from human blood (EDTA) or bone marrow (EDTA).

REAGENTS

Abbott RealTime IDH1 Control Kit

(List No. 08N90-080)

- Abbott RealTime IDH1 Negative Control**
(List No. 8N90Z) (5 vials, 0.210 mL per vial). <0.01% noninfectious plasmid DNA in a buffered solution with carrier DNA. Preservatives: sodium azide and 0.15% ProClin® 950.
- Abbott RealTime IDH1 Positive Control**
(List No. 8N90W) (5 vials, 0.210 mL per vial). <0.01% noninfectious plasmid DNA in a buffered solution with carrier DNA. Preservatives: sodium azide and 0.15% ProClin 950.

The Abbott RealTime IDH1 Control Kit must only be used with the Abbott RealTime IDH1 Amplification Reagent Kit (List No. 08N90-090 and 08N90-075).

WARNINGS AND PRECAUTIONS

In Vitro Diagnostic Medical Device

For In Vitro Diagnostic Use

- Consult Abbott RealTime IDH1 package insert instructions for use.
- Do not use beyond expiration date.
- The and vials are intended for single use only; remaining reagents after first use should be discarded.

The controls contain sodium azide and 2-methyl-2H-isothiazol-3-one, which is a component of ProClin 950.

Abbott RealTime IDH1 Negative Control

Abbott RealTime IDH1 Positive Control

The following warnings apply:

Warning	
	H317 May cause an allergic skin reaction.
	EUH032 Contact with acids liberates very toxic gas.
	P261 Avoid breathing mist/vapours/spray.
	P280 Wear protective gloves/protective clothing/eye protection.
	P272 Contaminated work clothing should not be allowed out of the workplace.
	P302+P352 IF ON SKIN: Wash with plenty of water.
	P333+P313 If skin irritation or rash occurs: Get medical advice/attention.
	P362+P364 Take off contaminated clothing and wash it before reuse.
	P501 Dispose of contents/container in accordance with local regulations.

Safety Data Sheet Statement: Important information regarding the safe handling, transport, and disposal of this product is contained in the Safety Data Sheet.

STORAGE INSTRUCTIONS

The Abbott RealTime IDH1 Control Kit must be stored at -25 to -15 °C.

SHIPPING CONDITIONS

The product is shipped on dry ice.

If assay reagents are received in a condition contrary to the label recommendation, or are damaged, contact Abbott Technical Services.

TECHNICAL ASSISTANCE

For technical assistance, call Abbott Molecular Technical Services at 1-800-553-7042 or visit the Abbott Molecular website at <http://www.abbottmolecular.com>.

Abbott RealTime is a trademark of Abbott Laboratories.

All other trademarks are property of their respective owners.

Abbott Molecular Inc.
1300 East Touhy Avenue
Des Plaines, IL 60018 USA



Abbott GmbH & Co. KG
Max-Planck-Ring 2
65205 Wiesbaden, Germany

© 2017, 2018 Abbott Laboratories

www.abbottmolecular.com
February 2018
51-608284/R2